

UV-Vis Spectrophotometer (Initial Specifications)

Wavelength range = 190nm-1100nm

Wavelength accuracy = $\pm 0.2\text{nm}$ - $\pm 1\text{nm}$

Spectral Bandwidth = 1nm - 2nm

Optical double beam

USB computer interface

Photometric Accuracy = $\pm 0.001A$ - $\pm 0.002A$


29/08/2010